5141 Schuylkill Street
Columbus, OH 43220-3235

WCL Corporation

Tel: 614-451-9256
Mobile 614-439-5528
bill.lowers@wclcorp.com

REPORT AND CERTIFICATION FOR THE EVALUATION OF A WHEELCHAIR SCALE MODEL DS9100-ADA MANUFACTURED BY DORAN SCALES, INC.

PRODUCT DESCRIPTION

The product being evaluated is an electronic scale that is intended to conform to the requirements of the applicable U.S. Access Board Medical Diagnostic Equipment Accessibility Standard *M303 Diagnostic Equipment Used by Patients Seated in a Wheelchair*.

CONFORMITY ASSESSMENT

Each element of the standards in question has been evaluated and their conformance has been documented in a Checklist. WCL Corporation personnel visited Doran Scales on October 27, 2026 to verify measurements and that the construction was conforming to drawings shown in the DOCUMENTATION section of this report. The sample evaluated at Doran Scales is a production sample used for marketing purposes and it had no label or serial number, but is identical to models that will be sold.

DOCUMENTATION

The documentation that has been used to support this conformity assessment are:

1. Drawing number File name - DS9100 ADA edrawing.easm
2. Instruction Document MAN0374 Rev 0
3. Bill of Material File name "DS9100ADA_BOM)translated.xlsx
4. Label Drawing LBL0330 Rev 00
5. Anti-Slip Tape file name – PVC_Emery_AntiSlip_Tape_Specification.pdf
6. Anti-Slip Pad file name – PVC_AntiSlip_Pad_Specification.pdf

WCL CORPORATION OPINION REGARDING CONFORMITY ASSESSMENT

WCL Corporation's opinion regarding the conformity assessment of the Doran Scales, Inc's. Model DS9100-ADA Wheelchair Scale is that it conforms to all elements of U.S. Access Board Medical Diagnostic Equipment Accessibility Standards *M303 Diagnostic Equipment Used by Patients Seated in a Wheelchair*. The assessment method included examination of 3D Design drawings, measurement of a production sample of the scale assembly, consultation with Doran Scales, Inc's. design engineering staff and observations by the assessor, Mr. William C. Lowers, President of WCL Corporation.

To summarize WCL Corporation's assessment, it is our opinion that the Model DS9100-ADA scale is fully conforming to all elements of the requirements of the applicable standards, and we Certify that result.

NOTE: WCL Corporation is not a Nationally Recognized Test Laboratory (NRTL). End of Report.

William C. Lowers



William C. Lowers
Electronic Signature Log
Number 236

President
WCL Corporation
DATE: November 9, 2026 Revision 01 November 17, 2026

5141 Schuylkill Street
Columbus, OH 43220-3235

WCL
Corporation

Tel: 614-451-9256
Mobile 614-439-5528
bill.lowers@wclcorp.com

Revision History

Revision 01 – November 16, 2026 – Removed strikethrough in report text indicating non-applicable parts of US Access Board Medical Diagnostic Equipment Accessibility Standard M303. This is an editorial change for clarity of reading, and no conclusions or opinions were changed. Added Letterhead and changed title to include the word "CERTIFICATION". The font size was reduced. The electronic signature log number 236 remains valid.